

**15/ 50 mL Centrifuge Tube
Magnetic Separation Rack
MSR1550**



Catalog # (P/N): MSR1550

Precautions:

CAUTION - STRONG MAGNETIC FIELD. This Permagen Labware Magnetic Separation Rack contains very strong neodymium iron boron (NdFeB) magnets. People with pacemakers or other forms of Implantable cardioverter-defibrillators (ICD's) should always avoid this product as it may interfere with its operation. Never allow this product to come into contact with other magnets, metal, magnetized media such as credit cards etc., as damage and or injury can occur. Do not Autoclave or heat above 80°C, doing so will damage the magnets.

USER INFORMATION

Product Description:

Permagen Labware's 15/ 50ml Magnetic Separation Rack is designed for large separations of common magnetic bead chemistries and protocols with up to (3) 50ml Tubes and or (3) 15ml Tubes. The magnets in our separation rack will pull beads to the inside wall of each well, leaving a clean area for removal.

Care/ Cleaning:

If solutions are spilled onto the separation rack carefully rinse with tap water and allow to air dry completely before use. A mild soap may also be used if more extreme cleaning is necessary. Always use care when storing product when not in use.

OTHER INFORMATION

Warranty and Liability:

For research use only. Not for use in diagnostic procedures.

Permagen Labware, LTD. Is focused on delivering industry leading product quality, and performance. Please visit www.permagenlabware.com for full warranty terms and conditions.

Any information contained herein is subject to change without notice. Permagen Labware, LTD. assumes no responsibility for any errors that may appear in this document.

Permagen Labware is a registered trademark of Permagen Labware, LTD
© 2015 Permagen Labware, LTD. All rights reserved.